

Aircheck Report and Certificate

From:
Trace Analytics, LLC
15768 Hamilton Pool Road
Austin, Texas 78738

800-247-1024 • 512-263-0000
Fax 512-263-0002
E-mail service@AirCheckLab.com

To:
Mr. Gerald Bayus, Jr.
RecTEC Divers LLC
101 Leckrone Way
Cortland, OH 44410

TRACE Analytics LLC

Report 19-16967, Sampled on

5/20/2019



Analysis Certificate

Next Sample Due Quarterly, Approximately

8/20/2019

RecTEC Divers LLC

IS IN COMPLIANCE WITH THE AIR/GAS QUALITY PORTION OF THE SPECIFICATION:
OXYGEN COMPAT. AIR-2003 (I) & CGA G-7.1-2011 GRADE E (2)

AS ANALYZED AND REPORTED ON THIS CERTIFICATE
FOR THE SAMPLE DESCRIBED UNDER SECTION "SAMPLE & REPORT INFORMATION"



American Assn for Laboratory Accreditation
1991: Certificate No. 322.01 Chemical Field of Testing

Maria Sandoval
Maria Sandoval, Laboratory Director

Analytical Test Methods		Media Sampled	Estimate of Uncertainty
Gases & Vapors	CAT-A-01 Gas Chromatography/Mass Spectrometry	Source Bottle: 790643	The average analytical uncertainty (k=2) is 98.8±2.4% (relative) at the specification limit for the ten compounds normally reported. For uncertainty information for a specific compound, contact Trace Analytics.
Oil & Particulate	CAT-A-03 Analytical Gravimetry	Ambient Bottle: 434773	
Particle Size	CAT-A-04 Optical Microscopy	Source Filter: 168981	
Pressure Dew Point	CAT-A-07 Gas Detector Tube	Detector Tube: Draeger 5-a/P	

Results relate only to items tested. This report shall not be reproduced except in full without the written permission of Trace Analytics, LLC
© Copyright 2019, Trace Analytics, LLC

Sample & Report Information

Results of Test: PASS

Sampled For	RecTEC Divers LLC	Analytes	Source Results	Ambient Results	Specification ¹ Allowable Limits
Sampled By	Gerald Bayus Jr.	Oxygen, Volume %	20.7	20.9	20-22
Sampled On	5/20/2019	Nitrogen, Volume %	78.4	78.2	N/A
Received On	5/23/2019	Argon, Volume %	0.9	0.9	N/A
Analyzed On	5/23/2019	Nitrogen Plus Argon, Volume %	79.3	N/A	N/A
Sampled From	Compressor	Carbon Monoxide (CO), ppmv	<0.5	<0.5	2
Make	Poseidon	Carbon Dioxide (CO ₂), ppmv	8	459	1000
Model	PFU150	Water Content (H ₂ O), ppmv/Dewpoint, °F	20.3 / -68	N/A	N/A / N/A (W)
Serial No.	PFU-15097042405	Atmospheric Dew Point, °F (DT)	-49	N/A	N/A
		TVHC (including CH ₄), ppmv	1.9	5.7	25
		Methane (CH ₄), ppmv	1.9	2.9	N/A
		TVHC (excluding CH ₄), ppmv	<0.7	2.8	N/A
		Oil (condensed) & Particulate, mg/m ³	0.04	N/A	0.1
Hours	2305	Odor (provided by customer)	None/Slight	N/A	None/Slight
Sample Phase	After Filter Change	Other	N/A	N/A	N/A
Customer Comments	Replaced Final Filter. Cleaned/Replaced All Sinterec Filters, Filter Tower And Moisture Separator. Outside Temp 66 Degrees Humidity 67%.	Other	N/A	N/A	N/A
		Other	N/A	N/A	N/A

PASS

(2) Water content is reported although not required by CGA G-7.1 for breathing air not used in conjunction with SCBA.

(I) This specification for oxygen compatible air is taken from ANSI/CGA G-7.1 Grade E as modified by International Association of Nitrox and Technical Divers (IANTD) in their document Blending Standards, 2003.

(W) Dew point is expressed in °F at one atmosphere pressure absolute.

(DT) Dew point is calculated at 1 atmosphere pressure (14.7 psia) from the detector tube reading.

Report Number 19-16967
Customer ID 32507
Date Reported 5/28/2019
Frequency Quarterly

Next Sample Due Approx. **8/20/2019**

We Do One Thing –Test Compressed Air

www.AirCheckLab.com

Aircheck Report and Certificate

From:
Trace Analytics, LLC
15768 Hamilton Pool Road
Austin, Texas 78738

800-247-1024 • 512-263-0000
Fax 512-263-0002
E-mail service@AirCheckLab.com

To:
Mr. Gerald Bayus, Jr.
RecTEC Divers LLC
101 Leckrone Way
Cortland, OH 44410

TRACE Analytics LLC



Analysis Certificate

Report 18-13705, Sampled on

5/2/2018

Next Sample Due Quarterly, Approximately

8/2/2018

RecTEC Divers LLC

IS IN COMPLIANCE WITH THE AIR/GAS QUALITY PORTION OF THE SPECIFICATION:
OXYGEN COMPAT. AIR-2003 (I) & CGA G-7.1-2011 GRADE E (2)
AS ANALYZED AND REPORTED ON THIS CERTIFICATE
FOR THE SAMPLE DESCRIBED UNDER SECTION "SAMPLE & REPORT INFORMATION"



American Assn for Laboratory Accreditation
1991: Certificate No. 322.01 Chemical Field of Testing

R. A. Smith
Richard A. Smith, Laboratory Director

Analytical Test Methods		Media Sampled	Estimate of Uncertainty
Gases & Vapors	CAT-A-01 Gas Chromatography/Mass Spectrometry	Source Bottle: 765995	The average analytical uncertainty (k=2) is 98.8±2.4% (relative) at the specification limit for the ten compounds normally reported. For uncertainty information for a specific compound, contact Trace Analytics.
Oil & Particulate	CAT-A-03 Analytical Gravimetry	Ambient Bottle: N/A	
Particle Size	CAT-A-04 Optical Microscopy	Source Filter: 3100	
Pressure Dew Point	CAT-A-07 Gas Detector Tube	Detector Tube: Draeger 5-a/P	

Results relate only to items tested. This report shall not be reproduced except in full without the written permission of Trace Analytics, LLC
© Copyright 2018, Trace Analytics, LLC

Sample & Report Information

Results of Test: PASS

Sampled For	RecTEC Divers LLC	Analytes	Source Results	Ambient Results	Specification ¹ Allowable Limits
Sampled By	Gerald Bayus Jr.	Oxygen, Volume %	21.0	N/A	20-22
Sampled On	5/2/2018	Nitrogen, Volume %	78.1	N/A	N/A
Received On	5/4/2018	Argon, Volume %	0.9	N/A	N/A
Analyzed On	5/4/2018	Nitrogen Plus Argon, Volume %	79.0	N/A	N/A
Sampled From	Compressor	Carbon Monoxide (CO), ppmv	1.0	N/A	2
Make	Poseidon	Carbon Dioxide (CO ₂), ppmv	516	N/A	1000
Model	PFU150	Water Content (H ₂ O), ppmv/Dewpoint, °F	207 / -32	N/A	N/A / N/A (W)
Serial No.	PFU-15097042405	Atmospheric Dew Point, °F (DT)	-49	N/A	N/A
		TVHC (including CH ₄), ppmv	4.5	N/A	25
		Methane (CH ₄), ppmv	2.5	N/A	N/A
		TVHC (excluding CH ₄), ppmv	2.0	N/A	N/A
		Oil (condensed) & Particulate, mg/m ³	0.04	N/A	0.1
Hours	2270	Odor (provided by customer)	None/Slight	N/A	None/Slight
Sample Phase	Routine	Other	N/A	N/A	N/A
Customer Comments		Other	N/A	N/A	N/A
		Other	N/A	N/A	N/A

(2) Water content is reported although not required by CGA G-7.1 for breathing air not used in conjunction with SCBA.

(I) This specification for oxygen compatible air is taken from ANSI/CGA G-7.1 Grade E as modified by International Association of Nitrox and Technical Divers (IANTD) in their document Blending Standards, 2003.

(W) Dew point is expressed in °F at one atmosphere pressure absolute.

(DT) Dew point is calculated at 1 atmosphere pressure (14.7 psia) from the detector tube reading.

Report Number 18-13705
Customer ID 32507
Date Reported 5/8/2018
Frequency Quarterly
Next Sample Due Approx. **8/2/2018**

PASS

Aircheck✓ Report and Certificate

From:
Trace Analytics, LLC
15768 Hamilton Pool Road
Austin, Texas 78738

800-247-1024 • 512-263-0000
Fax 512-263-0002
E-mail service@AirCheckLab.com

To:
Mr. Gerald Bayus, Jr.
RecTEC Divers LLC
101 Leckrone Way
Cortland, OH 44410

TRACEAnalytics^{LLC}

Report 17-14372, Sampled on

5/7/2017**Analysis Certificate**

Next Sample Due Quarterly, Approximately

8/7/2017**RecTEC Divers LLC**

IS IN COMPLIANCE WITH THE AIR/GAS QUALITY PORTION OF THE SPECIFICATION:
OXYGEN COMPAT. AIR-2003 (I) & CGA G-7.1-2011 GRADE E (2)
AS ANALYZED AND REPORTED ON THIS CERTIFICATE
FOR THE SAMPLE DESCRIBED UNDER SECTION "SAMPLE & REPORT INFORMATION"



American Assn for Laboratory Accreditation
1991: Certificate No. 322.01 Chemical Field of Testing

R. A. Smith
Richard A. Smith, Laboratory Director

Analytical Test Methods		Media Sampled	Estimate of Uncertainty
Gases & Vapors	CAT-A-01 Gas Chromatography/Mass Spectrometry	Source Bottle: 758212	The average analytical uncertainty (k=2) is 98.8±2.4% (relative) at the specification limit for the ten compounds normally reported. For uncertainty information for a specific compound, contact Trace Analytics.
Oil & Particulate	CAT-A-03 Analytical Gravimetry	Ambient Bottle: 433686	
Particle Size	CAT-A-04 Optical Microscopy	Source Filter: 3357	
Pressure Dew Point	CAT-A-07 Gas Detector Tube	Detector Tube: Draeger 5-a/P	

Results relate only to items tested. This report shall not be reproduced except in full without the written permission of Trace Analytics, LLC
© Copyright 2017, Trace Analytics, LLC

Sample & Report Information**Results of Test: PASS**

Sampled For	RecTEC Divers LLC	Analytes	Source Results	Ambient Results	Specification ¹ Allowable Limits
Sampled By	Gerald Bayus Jr.	Oxygen, Volume %	21.1	20.7	20-22
Sampled On	5/7/2017	Nitrogen, Volume %	78.0	78.4	N/A
Received On	5/12/2017	Argon, Volume %	0.9	0.9	N/A
Analyzed On	5/12/2017	Nitrogen Plus Argon, Volume %	78.9	N/A	N/A
Sampled From	Compressor	Carbon Monoxide (CO), ppmv	<0.3	<0.3	2
Make	Poseidon	Carbon Dioxide (CO ₂), ppmv	466	544	1000
Model	PFU150	Water Content (H ₂ O), ppmv/Dewpoint, °F	<3.4 / <-91	N/A	N/A / N/A (W)
Serial No.	PFU-15097042405	Atmospheric Dew Point, °F (DT)	-58	N/A	N/A
		TVHC (including CH ₄), ppmv	2.5	10.5	25
		Methane (CH ₄), ppmv	2.5	3.4	N/A
		TVHC (excluding CH ₄), ppmv	<0.7	7.1	N/A
		Oil (condensed) & Particulate, mg/m ³	<0.02	N/A	0.1
Hours	2234	Odor (provided by customer)	None/Slight	N/A	None/Slight
Sample Phase	Routine	Other	N/A	N/A	N/A
Customer Comments		Other	N/A	N/A	N/A
		Other	N/A	N/A	N/A

(2) Water content is reported although not required by CGA G-7.1 for breathing air not used in conjunction with SCBA.

(I) This specification for oxygen compatible air is taken from ANSI/CGA G-7.1 Grade E as modified by International Association of Nitrox and Technical Divers (IANTD) in their document Blending Standards, 2003.

(W) Dew point is expressed in °F at one atmosphere pressure absolute.

(DT) Dew point is calculated at 1 atmosphere pressure (14.7 psia) from the detector tube reading.

Report Number 17-14372
Customer ID 32507
Date Reported 5/15/2017
Frequency Quarterly

Next Sample Due Approx. **8/7/2017**

We Do One Thing – Test Compressed Air**www.AirCheckLab.com****PASS**

Aircheck Report and Certificate

From:
Trace Analytics, LLC
15768 Hamilton Pool Road
Austin, Texas 78738

800-247-1024 • 512-263-0000
Fax 512-263-0002
E-mail service@AirCheckLab.com

To:
Mr. Gerald Bayus, Jr.
RecTEC Divers LLC
101 Leckrone Way
Cortland, OH 44410

TRACEAnalytics^{LLC}

Report 16-13611, Sampled on

5/9/2016**Analysis Certificate**

Next Sample Due Quarterly, Approximately

8/9/2016**RecTEC Divers LLC**

IS IN COMPLIANCE WITH THE AIR/GAS QUALITY PORTION OF THE SPECIFICATION:
OXYGEN COMPAT. AIR-2003 (I) & CGA G-7.1-2011 GRADE E (2)
AS ANALYZED AND REPORTED ON THIS CERTIFICATE
FOR THE SAMPLE DESCRIBED UNDER SECTION "SAMPLE & REPORT INFORMATION"



American Assn for Laboratory Accreditation
1991: Certificate No. 322.01 Chemical Field of Testing

Richard A. Smith, Laboratory Director

Analytical Test Methods	Media Sampled	Estimate of Uncertainty
Gases & Vapors CAT-A-01 Gas Chromatography/Mass Spectrometry Oil & Particulate CAT-A-03 Analytical Gravimetry Particle Size CAT-A-04 Optical Microscopy	Source Bottle: 775703 Ambient Bottle: 433663 Source Filter: 2523	The average analytical uncertainty (k=2) is 98.8±2.4% (relative) at the specification limit for the ten compounds normally reported. For uncertainty information for a specific compound, contact Trace.

Results relate only to items tested. This report shall not be reproduced except in full without the written permission of Trace Analytics, LLC
© Copyright 2016, Trace Analytics, LLC

Sample & Report Information**Results of Test: PASS**

Sampled For	RecTEC Divers LLC	Analytes	Source Results	Ambient Results	Specification ¹ Allowable Limits
Sampled By	Gerald Bayus	Oxygen, Volume %	21.0	21.0	20-22
Sampled On	5/9/2016	Nitrogen, Volume %	78.1	78.2	N/A
Received On	5/13/2016	Argon, Volume %	0.9	0.8	N/A
Analyzed On	5/13/2016	Nitrogen Plus Argon, Volume %	79.0	N/A	N/A
Sampled From	Compressor	Carbon Monoxide (CO), ppmv	<0.3	<0.3	2
Make	Poseidon	Carbon Dioxide (CO ₂), ppmv	381	371	1000
Model	PFU150	Water Content (H ₂ O), ppmv/Dewpoint, °F	<3.4 / <-91	N/A	N/A / N/A (W)
Serial No.	PFU-15097042405	Atmospheric Dew Point, °F (DT)	-56	N/A	N/A
		TVHC (including CH ₄), ppmv	2.2	2.0	25
		Methane (CH ₄), ppmv	2.2	2.0	N/A
		TVHC (excluding CH ₄), ppmv	<0.7	<0.7	N/A
		Oil (condensed) & Particulate, mg/m ³	0.03	N/A	0.1
Hours	2218	Odor (provided by customer)	None/Slight	N/A	None/Slight
Sample Phase	Routine	Other	N/A	N/A	N/A
Customer Comments		Other	N/A	N/A	N/A
		Other	N/A	N/A	N/A

(2) Water content is reported although not required by CGA G-7.1 for breathing air not used in conjunction with SCBA.

(I) This specification for oxygen compatible air is taken from ANSI/CGA G-7.1 Grade E as modified by International Association of Nitrox and Technical Divers (IANTD) in their document Blending Standards, 2003.

(W) Dew point is expressed in °F at one atmosphere pressure absolute.

(DT) Dew point is calculated at 1 atmosphere pressure (14.7 psia) from the detector tube reading.

Report Number 16-13611
Customer ID 32507
Date Reported 5/16/2016
Frequency Quarterly
Next Sample Due Approx. **8/9/2016**

PASS

Aircheck Report and Certificate

From:
Trace Analytics, LLC
15768 Hamilton Pool Road
Austin, Texas 78738

800-247-1024 • 512-263-0000
Fax 512-263-0002
E-mail service@AirCheckLab.com

To:
Mr. Gerald Bayus, Jr.
RecTEC Divers LLC
101 Leckrone Way
Cortland, OH 44410

TRACE Analytics LLC

Report 15-28275, Sampled on

10/9/2015



Analysis Certificate

Next Sample Due Quarterly, Approximately

1/9/2016

RecTEC Divers LLC

IS IN COMPLIANCE WITH THE AIR/GAS QUALITY PORTION OF THE SPECIFICATION:
OXYGEN COMPAT. AIR-2003 (I) & CGA G-7.1-2011 GRADE E (2)

AS ANALYZED AND REPORTED ON THIS CERTIFICATE

FOR THE SAMPLE DESCRIBED UNDER SECTION "SAMPLE & REPORT INFORMATION"



American Assn for Laboratory Accreditation
1991: Certificate No. 322.01 Chemical Field of Testing

R. A. Smith
Richard A. Smith, Laboratory Director

Analytical Test Methods	Media Sampled	Estimate of Uncertainty
Gases & Vapors CAT-A-01 Gas Chromatography/Mass Spectrometry	Source Bottle: 758326	The average analytical uncertainty (k=2) is 98.8±2.4% (relative) at the specification limit for the ten compounds normally reported. For uncertainty information for a specific compound, contact Trace.
Oil & Particulate CAT-A-03 Analytical Gravimetry	Ambient Bottle: N/A	
Particle Size CAT-A-04 Optical Microscopy	Source Filter: 3502	

Results relate only to items tested. This report shall not be reproduced except in full without the written permission of Trace Analytics, LLC

© Copyright 2015, Trace Analytics, LLC

Sample & Report Information

Results of Test: PASS

Sampled For	RecTEC Divers LLC	Analytes	Source Results	Ambient Results	Specification ¹ Allowable Limits
Sampled By	Illegible	Oxygen, Volume %	20.8	N/A	20-22
Sampled On	10/9/2015	Nitrogen, Volume %	78.3	N/A	N/A
Received On	10/16/2015	Argon, Volume %	0.9	N/A	N/A
Analyzed On	10/16/2015	Nitrogen Plus Argon, Volume %	79.2	N/A	N/A
Sampled From	Compressor	Carbon Monoxide (CO), ppmv	<0.3	N/A	2
Make	Poseidon	Carbon Dioxide (CO ₂), ppmv	284	N/A	1000
Model	PFU150	Water Content (H ₂ O), ppmv/Dewpoint, °F	<3.4 / <-91	N/A	N/A / N/A (W)
Serial No.	PFU-15097042405	Atmospheric Dew Point, °F (DT)	-67	N/A	N/A
		TVHC (including CH ₄), ppmv	2.1	N/A	25
		Methane (CH ₄), ppmv	2.1	N/A	N/A
		TVHC (excluding CH ₄), ppmv	<0.7	N/A	N/A
		Oil (condensed) & Particulate, mg/m ³	<0.02	N/A	0.1
Hours	2202	Odor (provided by customer)	None/Slight	N/A	None/Slight
Sample Phase	Routine	Other	N/A	N/A	N/A
Customer Comments		Other	N/A	N/A	N/A
		Other	N/A	N/A	N/A

(2) Water content is reported although not required by CGA G-7.1 for breathing air not used in conjunction with SCBA.

(I) This specification for oxygen compatible air is taken from ANSI/CGA G-7.1 Grade E as modified by International Association of Nitrox and Technical Divers (IANTD) in their document Blending Standards, 2003.

(W) Dew point is expressed in °F at one atmosphere pressure absolute.

(DT) Dew point is calculated at 1 atmosphere pressure (14.7 psia) from the detector tube reading.

Report Number 15-28275
Customer ID 32507
Date Reported 10/19/2015
Frequency Quarterly

Next Sample Due Approx. **1/9/2016**

We Do One Thing –Test Compressed Air

www.AirCheckLab.com

PASS

Aircheck Report and Certificate

From:
Trace Analytics, LLC
15768 Hamilton Pool Road
Austin, Texas 78738

800-247-1024 • 512-263-0000
Fax 512-263-0002
E-mail service@AirCheckLab.com

To:
Mr. Gerald Bayus, Jr.
RecTEC Divers LLC
101 Leckrone Way
Cortland, OH 44410

TRACE Analytics LLC

Report 15-13454, Sampled on

5/13/2015



Analysis Certificate

Next Sample Due Quarterly, Approximately

8/13/2015

RecTEC Divers LLC

IS IN COMPLIANCE WITH THE AIR/GAS QUALITY PORTION OF THE SPECIFICATION:
OXYGEN COMPAT. AIR-2003 (I) & CGA G-7.1-2011 GRADE E (2)
AS ANALYZED AND REPORTED ON THIS CERTIFICATE
FOR THE SAMPLE DESCRIBED UNDER SECTION "SAMPLE & REPORT INFORMATION"



American Assn for Laboratory Accreditation
1991: Certificate No. 322.01 Chemical Field of Testing

R. A. Smith
Richard A. Smith, Laboratory Director

Analytical Test Methods	Media Sampled	Estimate of Uncertainty
Gases & Vapors CAT-A-01 Gas Chromatography/Mass Spectrometry Oil & Particulate CAT-A-03 Analytical Gravimetry Particle Size CAT-A-04 Optical Microscopy	Source Bottle: 714874 Ambient Bottle: 424522 Source Filter: 2201	The average analytical uncertainty (k=2) is 98.8±2.4% (relative) at the specification limit for the ten compounds normally reported. For uncertainty information for a specific compound, contact Trace.

Results relate only to items tested. This report shall not be reproduced except in full without the written permission of Trace Analytics, LLC
© Copyright 2015, Trace Analytics, LLC

Sample & Report Information

Results of Test: PASS

Sampled For	RecTEC Divers LLC	Analytes	Source Results	Ambient Results	Specification ¹ Allowable Limits
Sampled By	Gerald Bayus	Oxygen, Volume %	20.9	21.0	20-22
Sampled On	5/13/2015	Nitrogen, Volume %	78.2	78.1	N/A
Received On	5/18/2015	Argon, Volume %	0.9	0.9	N/A
Analyzed On	5/19/2015	Nitrogen Plus Argon, Volume %	79.1	N/A	N/A
Sampled From	Compressor	Carbon Monoxide (CO), ppmv	<0.3	<0.3	2
Make	Poseidon	Carbon Dioxide (CO ₂), ppmv	416	432	1000
Model	PFU150	Water Content (H ₂ O), ppmv/Dewpoint, °F	<3.4 / <-91	N/A	N/A / N/A (W)
Serial No.	PFU-15097042405	Atmospheric Dew Point, °F (DT)	-56	N/A	N/A
		TVHC (including CH ₄), ppmv	3.4	7.9	25
		Methane (CH ₄), ppmv	2.6	2.1	N/A
		TVHC (excluding CH ₄), ppmv	0.8	5.8	N/A
		Oil (condensed) & Particulate, mg/m ³	<0.02	N/A	0.1
Hours	2195	Odor (provided by customer)	None/Slight	N/A	None/Slight
Sample Phase	Routine	Other	N/A	N/A	N/A
Customer Comments		Other	N/A	N/A	N/A
		Other	N/A	N/A	N/A

(2) Water content is reported although not required by CGA G-7.1 for breathing air not used in conjunction with SCBA.

(I) This specification for oxygen compatible air is taken from ANSI/CGA G-7.1 Grade E as modified by International Association of Nitrox and Technical Divers (IANTD) in their document Blending Standards, 2003.

(W) Dew point is expressed in °F at one atmosphere pressure absolute.

(DT) Dew point is calculated at 1 atmosphere pressure (14.7 psia) from the detector tube reading.

Report Number 15-13454
Customer ID 32507
Date Reported 5/20/2015
Frequency Quarterly

Next Sample Due Approx. **8/13/2015**

We Do One Thing – Test Compressed Air

www.AirCheckLab.com

PASS

Aircheck Report and Certificate

From:
Trace Analytics, LLC
15768 Hamilton Pool Road
Austin, Texas 78738

800-247-1024 • 512-263-0000
Fax 512-263-0002
E-mail service@AirCheckLab.com

To:
Mr. Gerald Bayus, Jr.
RecTEC Divers LLC
101 Leckrone Way
Cortland, OH 44410

TRACE Analytics LLC

Report 14-26241, Sampled on

9/30/2014



Analysis Certificate

Next Sample Due Quarterly, Approximately

12/30/2014

RecTEC Divers LLC

IS IN COMPLIANCE WITH THE AIR/GAS QUALITY PORTION OF THE SPECIFICATION:
OXYGEN COMPAT. AIR-2003 (I) & CGA G-7.1-2011 GRADE E (2)
AS ANALYZED AND REPORTED ON THIS CERTIFICATE
FOR THE SAMPLE DESCRIBED UNDER SECTION "SAMPLE & REPORT INFORMATION"



American Assn for Laboratory Accreditation
1991: Certificate No. 322.01 Chemical Field of Testing

R.A. Smith
Richard A. Smith, Laboratory Director

Analytical Test Methods	Media Sampled	Estimate of Uncertainty
Gases & Vapors CAT-A-01 Gas Chromatography/Mass Spectrometry Oil & Particulate CAT-A-03 Analytical Gravimetry Particle Size CAT-A-04 Optical Microscopy	Source Bottle: 766244 Ambient Bottle: 429338 Source Filter: 3201	The average analytical uncertainty (k=2) is 98.8±2.4% (relative) at the specification limit for the ten compounds normally reported. For uncertainty information for a specific compound, contact Trace.

Results relate only to items tested. This report shall not be reproduced except in full without the written permission of Trace Analytics, LLC

© Copyright 2014, Trace Analytics, LLC

Sample & Report Information

Results of Test: PASS

Sampled For	RecTEC Divers LLC	Analytes	Source Results	Ambient Results	Specification ¹ Allowable Limits
Sampled By	Gerald Bayus Jr.	Oxygen, Volume %	20.9	21.0	20-22
Sampled On	9/30/2014	Nitrogen, Volume %	78.2	78.2	N/A
Received On	10/3/2014	Argon, Volume %	0.9	0.8	N/A
Analyzed On	10/3/2014	Nitrogen Plus Argon, Volume %	79.1	N/A	N/A
Sampled From	Compressor	Carbon Monoxide (CO), ppmv	<0.3	<0.3	2
Make	Poseidon	Carbon Dioxide (CO ₂), ppmv	360	383	1000
Model	PFU150	Water Content (H ₂ O), ppmv/Dewpoint, °F	23.5 / -66	N/A	N/A / N/A (W)
Serial No.	PFU-15097042405	Atmospheric Dew Point, °F (DT)	-87	N/A	N/A
		TVHC (including CH ₄), ppmv	2.5	2.5	25
		Methane (CH ₄), ppmv	2.5	2.5	N/A
		TVHC (excluding CH ₄), ppmv	<0.7	<0.7	N/A
		Oil (condensed) & Particulate, mg/m ³	0.04	N/A	0.1
Hours	2195	Odor (provided by customer)	None/Slight	N/A	None/Slight
Sample Phase	Routine	Other	N/A	N/A	N/A
Customer Comments		Other	N/A	N/A	N/A
		Other	N/A	N/A	N/A

(2) Water content is reported although not required by CGA G-7.1 for breathing air not used in conjunction with SCBA.

(1) This specification for oxygen compatible air is taken from ANSI/CGA G-7.1 Grade E as modified by International Association of Nitrox and Technical Divers (IANTD) in their document Blending Standards, 2003.

(W) Dew point is expressed in °F at one atmosphere pressure absolute.

Report Number 14-26241
Customer ID 32507
Date Reported 10/6/2014
Frequency Quarterly
Next Sample Due Approx. **12/30/2014**

PASS

Aircheck Report and Certificate

From:
Trace Analytics, LLC
15768 Hamilton Pool Road
Austin, Texas 78738

800-247-1024 • 512-263-0000
Fax 512-263-0002
E-mail service@AirCheckLab.com

To:
Mr. Gerald Bayus, Jr.
RecTEC Divers LLC
101 Leckrone Way
Cortland, OH 44410

TRACEAnalytics^{LLC}

Report 14-13082, Sampled on

5/16/2014**Analysis Certificate**

Next Sample Due Quarterly, Approximately

8/16/2014**RecTEC Divers LLC**

IS IN COMPLIANCE WITH THE AIR/GAS QUALITY PORTION OF THE SPECIFICATION:
OXYGEN COMPAT. AIR-2003 (I) & CGA G-7.1-2011 GRADE E (2)
AS ANALYZED AND REPORTED ON THIS CERTIFICATE
FOR THE SAMPLE DESCRIBED UNDER SECTION "SAMPLE & REPORT INFORMATION"



American Assn for Laboratory Accreditation
1991: Certificate No. 322.01 Chemical Field of Testing

R. A. Smith
Richard A. Smith, Laboratory Director

Analytical Test Methods	Media Sampled	Estimate of Uncertainty
Gases & Vapors CAT-A-01 Gas Chromatography/Mass Spectrometry	Source Bottle: 762814	The average analytical uncertainty (k=2) is 98.8±2.4% (relative) at the specification limit for the ten compounds normally reported. For uncertainty information for a specific compound, contact Trace.
Oil & Particulate CAT-A-03 Analytical Gravimetry	Ambient Bottle: 430098	
Particle Size CAT-A-04 Optical Microscopy	Source Filter: 2153	

Results relate only to items tested. This report shall not be reproduced except in full without the written permission of Trace Analytics, LLC

© Copyright 2014, Trace Analytics, LLC

Sample & Report Information**Results of Test: PASS**

Sampled For	RecTEC Divers LLC	Analytes	Source Results	Ambient Results	Specification ¹ Allowable Limits
Sampled By	Gerald Bayus Jr.	Oxygen, Volume %	21.0	20.9	20-22
Sampled On	5/16/2014	Nitrogen, Volume %	78.1	78.2	N/A
Received On	5/19/2014	Argon, Volume %	0.9	0.9	N/A
Analyzed On	5/19/2014	Nitrogen Plus Argon, Volume %	79.0	N/A	N/A
Sampled From	Compressor	Carbon Monoxide (CO), ppmv	<0.3	<0.3	2
Make	Poseidon	Carbon Dioxide (CO ₂), ppmv	475	524	1000
Model	PFU150	Water Content (H ₂ O), ppmv/Dewpoint, °F	15.3 / -72	N/A	N/A / N/A (W)
Serial No.	PFU-15097042405	Atmospheric Dew Point, °F (DT)	-58	N/A	N/A
		TVHC (including CH ₄), ppmv	2.3	2.0	25
		Methane (CH ₄), ppmv	2.3	2.0	N/A
		TVHC (excluding CH ₄), ppmv	<0.7	<0.7	N/A
		Oil (condensed) & Particulate, mg/m ³	0.02	N/A	0.1
Hours	2173	Odor (provided by customer)	None/Slight	N/A	None/Slight
Sample Phase	Routine	Other	N/A	N/A	N/A
Customer Comments		Other	N/A	N/A	N/A
		Other	N/A	N/A	N/A

(2) Water content is reported although not required by CGA G-7.1 for breathing air not used in conjunction with SCBA.

(I) This specification for oxygen compatible air is taken from ANSI/CGA G-7.1 Grade E as modified by International Association of Nitrox and Technical Divers (IANTD) in their document Blending Standards, 2003.

(W) Dew point is expressed in °F at one atmosphere pressure absolute.

Report Number 14-13082
Customer ID 32507
Date Reported 5/20/2014
Frequency Quarterly

Next Sample Due Approx. **8/16/2014**

We Do One Thing –Test Compressed Air

www.AirCheckLab.com

PASS

Aircheck Report and Certificate

From:
Trace Analytics, LLC
15768 Hamilton Pool Road
Austin, Texas 78738

800-247-1024 • 512-263-0000
Fax 512-263-0002
E-mail service@AirCheckLab.com

To:
Mr. Gerald Bayus, Jr.
RecTEC Divers LLC
101 Leckrone Way
Cortland, OH 44410

TRACE Analytics LLC

Report 13-32384, Sampled on

12/17/2013



Analysis Certificate

Next Sample Due Quarterly, Approximately

3/17/2014

RecTEC Divers LLC

IS IN COMPLIANCE WITH THE AIR/GAS QUALITY PORTION OF THE SPECIFICATION:
OXYGEN COMPAT. AIR-2003 (I) & CGA G-7.1-2011 GRADE E (2)
AS ANALYZED AND REPORTED ON THIS CERTIFICATE
FOR THE SAMPLE DESCRIBED UNDER SECTION "SAMPLE & REPORT INFORMATION"



American Assn for Laboratory Accreditation
1991: Certificate No. 322.01 Chemical Field of Testing

R.A. Smith
Richard A. Smith, Laboratory Director

Analytical Test Methods	Media Sampled	Estimate of Uncertainty
Gases & Vapors CAT-A-01 Gas Chromatography/Mass Spectrometry Oil & Particulate CAT-A-03 Analytical Gravimetry Particle Size CAT-A-04 Optical Microscopy	Source Bottle: 737158 Ambient Bottle: N/A Source Filter: 148473	The average analytical uncertainty (k=2) is 98.8±2.4% (relative) at the specification limit for the ten compounds normally reported. For uncertainty information for a specific compound, contact Trace.

Results relate only to items tested. This report shall not be reproduced except in full without the written permission of Trace Analytics, LLC

© Copyright 2013, Trace Analytics, LLC

Sample & Report Information

Results of Test: PASS

Sampled For	RecTEC Divers LLC	Analytes	Source Results	Ambient Results	Specification ¹ Allowable Limits
Sampled By	Gerald Bayus Jr.	Oxygen, Volume %	20.9	N/A	20-22
Sampled On	12/17/2013	Nitrogen, Volume %	78.3	N/A	N/A
Received On	12/23/2013	Argon, Volume %	0.8	N/A	N/A
Analyzed On	12/23/2013	Nitrogen Plus Argon, Volume %	79.1	N/A	N/A
Sampled From	Compressor	Carbon Monoxide (CO), ppmv	<0.3	N/A	2
Make	Poseidon	Carbon Dioxide (CO ₂), ppmv	14	N/A	1000
Model	PFU150	Water Content (H ₂ O), ppmv/Dewpoint, °F	<3.4 / <91	N/A	N/A / N/A (W)
Serial No.	PFU-15097042405	Atmospheric Dew Point, °F (DT)	-81	N/A	N/A
		TVHC (including CH ₄), ppmv	5.1	N/A	25
		Methane (CH ₄), ppmv	2.7	N/A	N/A
		TVHC (excluding CH ₄), ppmv	2.4	N/A	N/A
		Oil (condensed) & Particulate, mg/m ³	<0.04 (IAV)	N/A	0.1
Hours	2167	Odor (provided by customer)	None/Slight	N/A	None/Slight
Sample Phase	After Filter Change	Other	N/A	N/A	N/A
Customer Comments	All Filters, Air Filters, Separation Filters were changed at 2166 HRS.	Other	N/A	N/A	N/A
		Other	N/A	N/A	N/A

(2) Water content is reported although not required by CGA G-7.1 for breathing air not used in conjunction with SCBA.

(I) This specification for oxygen compatible air is taken from ANSI/CGA G-7.1 Grade E as modified by International Association of Nitrox and Technical Divers (IANTD) in their document Blending Standards, 2003.

(W) Dew point is expressed in °F at one atmosphere pressure absolute.

(DT) Dew point is calculated from the detector tube reading.

(IAV) To ensure accuracy, the air volume should be sufficient to provide a limit of quantitation of <1/4 of the specification limit; in this case 1000 L. Please refer to sampling instructions or contact Customer Service.

PASS

Aircheck Report and Certificate

From:
Trace Analytics, LLC
15768 Hamilton Pool Road
Austin, Texas 78738

800-247-1024 • 512-263-0000
Fax 512-263-0002
E-mail service@AirCheckLab.com

To:
Mr. Gerald Bayus, Jr.
RecTEC Divers LLC
101 Leckrone Way
Cortland, OH 44410

TRACE Analytics LLC

Report 13-32385, Sampled on

12/8/2013



Analysis Certificate

Next Sample Due Quarterly, Approximately

3/8/2014

RecTEC Divers LLC

IS IN COMPLIANCE WITH THE AIR/GAS QUALITY PORTION OF THE SPECIFICATION:
OXYGEN COMPAT. AIR-2003 (I) & CGA G-7.1-2011 GRADE E (2)
AS ANALYZED AND REPORTED ON THIS CERTIFICATE
FOR THE SAMPLE DESCRIBED UNDER SECTION "SAMPLE & REPORT INFORMATION"



American Assn for Laboratory Accreditation
1991: Certificate No. 322.01 Chemical Field of Testing

R. A. Smith
Richard A. Smith, Laboratory Director

Analytical Test Methods	Media Sampled	Estimate of Uncertainty
Gases & Vapors CAT-A-01 Gas Chromatography/Mass Spectrometry Oil & Particulate CAT-A-03 Analytical Gravimetry Particle Size CAT-A-04 Optical Microscopy	Source Bottle: 729202 Ambient Bottle: 428239 Source Filter: 1968	The average analytical uncertainty (k=2) is 98.8±2.4% (relative) at the specification limit for the ten compounds normally reported. For uncertainty information for a specific compound, contact Trace.

Results relate only to items tested. This report shall not be reproduced except in full without the written permission of Trace Analytics, LLC

© Copyright 2013, Trace Analytics, LLC

Sample & Report Information

Results of Test: PASS

Sampled For	RecTEC Divers LLC	Analytes	Source Results	Ambient Results	Specification ¹ Allowable Limits
Sampled By	Gerald Bayus Jr.	Oxygen, Volume %	20.7	20.9	20-22
Sampled On	12/8/2013	Nitrogen, Volume %	78.4	78.2	N/A
Received On	12/23/2013	Argon, Volume %	0.9	0.9	N/A
Analyzed On	12/23/2013	Nitrogen Plus Argon, Volume %	79.3	N/A	N/A
Sampled From	Compressor	Carbon Monoxide (CO), ppmv	<0.3	<0.3	2
Make	Poseidon	Carbon Dioxide (CO ₂), ppmv	<4	1117	1000
Model	PFU150	Water Content (H ₂ O), ppmv/Dewpoint, °F	<3.4 / <91	N/A	N/A / N/A (W)
Serial No.	PFU-15097042405	Atmospheric Dew Point, °F (DT)	-67	N/A	N/A
		TVHC (including CH ₄), ppmv	4.7	11.8	25
		Methane (CH ₄), ppmv	2.7	3.6	N/A
		TVHC (excluding CH ₄), ppmv	2.0	8.2	N/A
		Oil (condensed) & Particulate, mg/m ³	<0.05 (IAV)	N/A	0.1
Hours	2166	Odor (provided by customer)	None/Slight	N/A	None/Slight
Sample Phase	After Filter Change	Other	N/A	N/A	N/A
Customer Comments		Other	N/A	N/A	N/A
		Other	N/A	N/A	N/A

(2) Water content is reported although not required by CGA G-7.1 for breathing air not used in conjunction with SCBA.

(I) This specification for oxygen compatible air is taken from ANSI/CGA G-7.1 Grade E as modified by International Association of Nitrox and Technical Divers (IANTD) in their document Blending Standards, 2003.

(W) Dew point is expressed in °F at one atmosphere pressure absolute.

(DT) Dew point is calculated from the detector tube reading.

(IAV) To ensure accuracy, the air volume should be sufficient to provide a limit of quantitation of <1/4 of the specification limit; in this case 1000 L. Please refer to sampling instructions or contact Customer Service.

PASS